



# CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

**Montana Department of Justice  
Forensic Science Division – Breath Alcohol Section  
2679 Palmer Street, Missoula, Montana 59808 USA**

Fulfills the requirements of

**ISO/IEC 17025:2017  
ANAB Forensic Testing & Calibration AR 3125:2019**

In the field of

**Forensic Calibration**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

Pamela L. Sale, Vice President, Forensics

Expiry Date: 30 November 2023  
Certificate Number: FC-0018





ANSI National Accreditation Board

SCOPE OF ACCREDITATION TO:
ISO/IEC 17025:2017
ANAB Forensic Testing & Calibration AR 3125:2019

Montana Department of Justice
Forensic Science Division
Breath Alcohol Section

2679 Palmer Street
Missoula, Montana 59808 USA

FORENSIC CALIBRATION

Expiry Date: 30 November 2023

Certificate Number: FC-0018

Discipline – Toxicology

Table with 5 columns: Component Calibrated, Calibration Method, Breath Alcohol Instrument Reporting Range, Expanded Uncertainty of Measurement, and Metrological Traceability. Row 1: Infrared Spectroscopy, BA-200, 0.020-0.300 g ethanol/210 L vapor, ± 0.0069 g ethanol/210L vapor 95.45%, k=2, Certified Reference Material.

Note 1: Component calibrated indicates the breath alcohol measuring instrument detector(s) used for measuring breath alcohol. Detector(s) used for qualitative purposes will not be listed on the Scope of Accreditation.

Note 2: The expanded uncertainty is expressed at the stated coverage probability (e.g., 95.45%) and coverage factor (e.g., k = 2). The actual measurement uncertainty reported on a certificate of calibration will be equal to or greater than the expanded uncertainty of measurement stated on this scope.

Handwritten signature of Pamela L. Sale

Pamela L. Sale
Vice President, Forensics